

UNITED STATES ENVIRONMENTAL PROTECTION AGENCY  
NATIONAL EXPOSURE RESEARCH LABORATORY  
EXPOSURE METHODS AND MEASUREMENT DIVISION (MD-D205-03)  
Research Triangle Park, NC 27711

**LIST OF DESIGNATED REFERENCE AND EQUIVALENT METHODS**

**Issue Date: June 15, 2018**  
([www.epa.gov/ttn/amtic/criteria.html](http://www.epa.gov/ttn/amtic/criteria.html))

***Horiba Model APSA-370 Ambient SO<sub>2</sub> Monitor***

**Automated Equivalent Method: EQSA-0506-159**

“Horiba Instruments Incorporated Model APSA-370 Ambient SO<sub>2</sub> Monitor,” operated with a full scale fixed measurement range of 0 - 0.50 ppm, with the automatic range switching off, at any environmental temperature in the range of 20°C to 30°C.<sup>2</sup>

***Federal Register: Vol. 71, page 25587, 05/01/2006***